K033524 P/3

DEC 1 8 2003

## 11 510(k) SUMMARY

#### 11.0 510(k) Summary

Coapt Systems is providing a summary of the safety and effectiveness information available for the ENDOTINE Chin<sup>TM</sup> 3.5 Device. This 510(k) summary of safety and effectiveness information is submitted in accordance with the requirements of 21 CFR §807.92 and pursuant to Section 12, Part (a)(i)(3A) of the Safe Medical Devices Act of 1990.

## SPONSOR/APPLICANT NAME AND ADDRESS

Coapt Systems, Inc. 1820 Embarcadero Road Palo Alto, CA

Telephone: (650) 461-7600 Facsimile: (650) 213-9336

## **CONTACT INFORMATION**

Lori DonDiego
Director, Regulatory Affairs
Coapt Systems, Inc.
1820 Embarcadero Road
Palo Alto, CA

Telephone: (650) 461-7647 Facsimile: (650) 213-9336

Email: ldondiego@coaptsystems.com

## DATE OF PREPARATION OF 510(K) SUMMARY

November 6, 2003

#### DEVICE TRADE OR PROPRIETARY NAME

ENDOTINE Chin<sup>TM</sup> 3.5 Device

### DEVICE COMMON OR CLASSIFICATION NAME

Classification Name: Smooth or threaded metallic bone fixation fastener

Regulation Number: 888.3040

Class:

П

Product Code:

**HWC** 

# IDENTIFICATION OF THE LEGALLY MARKETED DEVICES TO WHICH **EQUIVALENCE IS BEING CLAIMED**

Name of Predicate Device	Name of Manufacturer	510(k) or PMA Number
ENDOTINE Forehead™ Device	Coapt Systems, Inc	K014153/K023992
Lactosorb® Panels	Walter Lorenz Surgical	K974309

#### DEVICE DESCRIPTION

The ENDOTINE Chin™ 3.5 Device kit consists of the chin implant with insertion tool and a drill bit. The Chin device along with its insertion tool and drill bit are supplied sterile for single use only.

#### INTENDED USE STATEMENT

The ENDOTINE Chin<sup>TM</sup> 3.5 is intended for use in subperiosteal chinplasty surgery. The ENDOTINE Chin<sup>TM</sup> 3.5 is specifically indicated for use to fixate the subdermis to the mentum in chinplasty.

## SUBSTANTIAL EQUIVLANCE COMPARISON

#### 1. Indications Summary

The ENDOTINE Forehead<sup>TM</sup> Device intended use statement is nearly identical to the the new ENDOTINE Chin Device intended use. Both ENDOTINE devices (Forehead and Chin) are intended to elevate and fixate soft tissue in facial surgeries. Moreover, the anchoring methods are very similar for both devices, applying a pulling force to the skeletal system for support. Both devices are intended for soft tissue suspension using the patented multi-point technology developed by Coapt Systems.

The different anatomical sites for use between the ENDOTINE Chin™ 3.5 and ENDOTINE Forehead™ Device do not suggest new issues of safety and effectiveness when used in the chin region that could affect the substantial equivalence determination.

#### 2. Technological Characteristics Summary

The ENDOTINE Chin™ 3.5 is substantially equivalent in design, materials and fundamental scientific technology to the ENDOTINE Forehead and Lactosorb Panel predicate devices. Further, the technological characteristics of the ENDOTINE Chin 3.5 are similar to many absorbable, implantable general, orthopedic and plastic surgery devices legally distributed by device manufacturers. Any differences between the ENDOTINE Chin<sup>TM</sup> 3.5 and the predicate devices are minor and do not raise new issues regarding safety or effectiveness. This statement is substantiated by the clinical history to

13/3

date for the ENDOTINE Forehead Device and an established safety profile of the device material in both the forehead and chin regions.

#### 3. Performance Summary

The ENDOTINE Chin<sup>TM</sup> 3.5 Device is safe and appropriate for the intended use due to the following:

- Its similarity to the predicate device, the ENDOTINE Forehead Device.
- Design performance studies that included cadaveric modeling and evaluation, and strength testing results that exceeded specifications.
- Feedback and user observation from several leading surgeons.

The ENDOTINE Chin™ 3.5 performance data meet the applicable standards and fulfill the device requirements as defined in the design specifications.

An appropriate and complete testing program supports the ENDOTINE Chin™ 3.5 is suitable to perform and operate as clinically intended.

## SUBSTANTIAL EQUIVALENCE CONCLUSION

Based on the design, materials, function, intended use, and performance evaluations discussed herein, Coapt Systems believes the ENDOTINE Chin<sup>TM</sup> 3.5 Device is substantially equivalent to the selected predicate devices currently marketed under the Federal Food, Drug and Cosmetic Act. No new issues of safety or effectiveness were raised for the ENDOTINE Chin<sup>TM</sup> 3.5 Device. Therefore, safety and effectiveness are reasonably assured, justifying 510(k) clearance for commercial sale.





Food and Drug Administration 9200 Corporate Boulevard Rockville MD 20850

DEC 1 8 2003

Ms. Lori DonDiego Director, Regulatory Affairs Coapt Systems, Inc. 1820 Embarcadero Road Palo Alto, California 94303

Re: K033524

Trade/Device Name: ENDOTINE Chin™ 3.5

Regulation Number: 21 CFR 888.3040; 21 CFR 878.4493

Regulation Name: Smooth or threaded metallic bone fixation fastener; Absorbable

poly(glycolide/L-lactide) surgical suture

Regulatory Class: II

Product Code: HWC, GAM Dated: November 6, 2003 Received: November 7, 2003

### Dear Ms. DonDiego:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug. and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the <u>Federal Register</u>.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (301) 594-4659. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <a href="http://www.fda.gov/cdrh/dsma/dsmamain.html">http://www.fda.gov/cdrh/dsma/dsmamain.html</a>

Sincerely yours,

for Celia M. Witten, Ph.D., M.D.

Director

Division of General, Restorative and Neurological Devices Office of Device Evaluation Center for Devices and Radiological Health

Miriam C. Provost

Enclosure

510(k) Number:

# 4 STATEMENT OF INDICATIONS FOR USE

K033524

Device Name:	ENDOTINE Chin™ 3.5 Device	
Indications for Use:	The ENDOTINE Chin <sup>TM</sup> 3.5 is intended for use in subperiosteal chinplasty surgery. The ENDOTINE Chin <sup>TM</sup> 3.5 is specifically indicated for use to fixate the subdermis to the mentum in chinplasty.	
Prescription Use X (Part 21 CFR 801 Subpart D)	AND/OR	Over-The-Counter Use(21 CFR 807 Subpart C)
(PLEASE DO NOT WRIT IF NEEDED)	E BELOW THIS LINE-CO	NTINUE ON ANOTHER PAGE
Concurre	nce of CDRH, Office of Dev	ice Evaluation (ODE)
Division and Neu	uam C. Provost in Sign-Off) in of General, Restorative prological Devices	
***.3(k)	K033524	erratiid